

Canister Cleaning Certificate

Analyst	BHARATH.R.KATHI
Run Time, Date	03/18/2010
System	Entech Can Cleaning System,Model-3100
Gas	UHP Nitrogen
Batch Number	CC031810
Number of Canisters Batched	9
Canister Number	1350,4569,4570,1197,1192,1186,1194,1189,1188
Final Vac (mtorr)	50
Canister for analysis and Certification	1189
TNMOC of the Certified Canister	<20 ppb,V
Certified by GC/MS-Analyst Date File:	KBR C:\HPCHEM\1\DATA\P1003\18\CC031810
Date Canisters- Releasing	03/23/10
Log Book #	9G03-216

Comments: CITY OF HOUSTON CANISTER NUMBERS: 1350, 4569, 4570

Analyst: 

Date: 3/19/10

Reviewed by: 

Date: 3/19/10

Released by: 

Date 3/23/10